



BioPro Fasciotome™

Indications for use:

Percutaneous partial release of the plantar fascia.

Sterile:

Sterilized with ethylene oxide gas. Caution: For one procedure only. Do not re-sterilize. Do not use if package is open or damaged. This is a single use device. Re-use of this device can result in the transfer of materials not limited to bone, tissue, blood, or infectious disease. The device is provided sterile and re-sterilization of the device has not been validated.

Instructions for use:

- 1. After an appropriate local block has been obtained, locate the medial band of the plantar fascia by dorsiflexing the foot and great toe. This allows you to palpate the medial band in the arch area and follow it back to proximal of the heel.
2. The percutaneous portal is located by bisecting the medial malleolus and drawing the bisection line down to the medial plantar area of the foot. This approach brings you to a safe location for percutaneous release of the fascia by being proximal to the medial and lateral plantar nerves.
3. Ultrasound imaging can aid in the accurate placement of the percutaneous Fasciotomy.
4. The foot should be appropriately prepped and draped. Utilizing the Fasciotome™ the skin is penetrated and the Fasciotome™ is brought in contact with the medial band of the plantar fascia, cutting the medial band of the plantar fascia and a portion of the central band, if desired.
5. After the Fasciotome™ releases the medial band, there should be a loss of the prominence of the medial band to the fascia in the arch area.
6. No suturing of the portal is necessary, dry sterile dressing, and a modified Campbell's rest strapping of the foot and ankle is recommended.
7. Patient may immediately weight bear.

BioPro, Inc.
2929 Lapeer Road
Port Huron, MI 48060
810-982-7777 (P)
810-982-7794 (F)

BP060713 01



BioPro Fasciotome™

Indications for use:

Percutaneous partial release of the plantar fascia.

Sterile:

Sterilized with ethylene oxide gas. Caution: For one procedure only. Do not re-sterilize. Do not use if package is open or damaged. This is a single use device. Re-use of this device can result in the transfer of materials not limited to bone, tissue, blood, or infectious disease. The device is provided sterile and re-sterilization of the device has not been validated.

Instructions for use:

- 1. After an appropriate local block has been obtained, locate the medial band of the plantar fascia by dorsiflexing the foot and great toe. This allows you to palpate the medial band in the arch area and follow it back to proximal of the heel.
2. The percutaneous portal is located by bisecting the medial malleolus and drawing the bisection line down to the medial plantar area of the foot. This approach brings you to a safe location for percutaneous release of the fascia by being proximal to the medial and lateral plantar nerves.
3. Ultrasound imaging can aid in the accurate placement of the percutaneous Fasciotomy.
4. The foot should be appropriately prepped and draped. Utilizing the Fasciotome™ the skin is penetrated and the Fasciotome™ is brought in contact with the medial band of the plantar fascia, cutting the medial band of the plantar fascia and a portion of the central band, if desired.
5. After the Fasciotome™ releases the medial band, there should be a loss of the prominence of the medial band to the fascia in the arch area.
6. No suturing of the portal is necessary, dry sterile dressing, and a modified Campbell's rest strapping of the foot and ankle is recommended.
7. Patient may immediately weight bear.

BioPro, Inc.
2929 Lapeer Road
Port Huron, MI 48060
810-982-7777 (P)
810-982-7794 (F)

BP060713 01



BioPro Fasciotome™

Indications for use:

Percutaneous partial release of the plantar fascia.

Sterile:

Sterilized with ethylene oxide gas. Caution: For one procedure only. Do not re-sterilize. Do not use if package is open or damaged. This is a single use device. Re-use of this device can result in the transfer of materials not limited to bone, tissue, blood, or infectious disease. The device is provided sterile and re-sterilization of the device has not been validated.

Instructions for use:

- 1. After an appropriate local block has been obtained, locate the medial band of the plantar fascia by dorsiflexing the foot and great toe. This allows you to palpate the medial band in the arch area and follow it back to proximal of the heel.
2. The percutaneous portal is located by bisecting the medial malleolus and drawing the bisection line down to the medial plantar area of the foot. This approach brings you to a safe location for percutaneous release of the fascia by being proximal to the medial and lateral plantar nerves.
3. Ultrasound imaging can aid in the accurate placement of the percutaneous Fasciotomy.
4. The foot should be appropriately prepped and draped. Utilizing the Fasciotome™ the skin is penetrated and the Fasciotome™ is brought in contact with the medial band of the plantar fascia, cutting the medial band of the plantar fascia and a portion of the central band, if desired.
5. After the Fasciotome™ releases the medial band, there should be a loss of the prominence of the medial band to the fascia in the arch area.
6. No suturing of the portal is necessary, dry sterile dressing, and a modified Campbell's rest strapping of the foot and ankle is recommended.
7. Patient may immediately weight bear.

BioPro, Inc.
2929 Lapeer Road
Port Huron, MI 48060
810-982-7777 (P)
810-982-7794 (F)

BP060713 01



BioPro Fasciotome™

Indications for use:

Percutaneous partial release of the plantar fascia.

Sterile:

Sterilized with ethylene oxide gas. Caution: For one procedure only. Do not re-sterilize. Do not use if package is open or damaged. This is a single use device. Re-use of this device can result in the transfer of materials not limited to bone, tissue, blood, or infectious disease. The device is provided sterile and re-sterilization of the device has not been validated.

Instructions for use:

- 1. After an appropriate local block has been obtained, locate the medial band of the plantar fascia by dorsiflexing the foot and great toe. This allows you to palpate the medial band in the arch area and follow it back to proximal of the heel.
2. The percutaneous portal is located by bisecting the medial malleolus and drawing the bisection line down to the medial plantar area of the foot. This approach brings you to a safe location for percutaneous release of the fascia by being proximal to the medial and lateral plantar nerves.
3. Ultrasound imaging can aid in the accurate placement of the percutaneous Fasciotomy.
4. The foot should be appropriately prepped and draped. Utilizing the Fasciotome™ the skin is penetrated and the Fasciotome™ is brought in contact with the medial band of the plantar fascia, cutting the medial band of the plantar fascia and a portion of the central band, if desired.
5. After the Fasciotome™ releases the medial band, there should be a loss of the prominence of the medial band to the fascia in the arch area.
6. No suturing of the portal is necessary, dry sterile dressing, and a modified Campbell's rest strapping of the foot and ankle is recommended.
7. Patient may immediately weight bear.

BioPro, Inc.
2929 Lapeer Road
Port Huron, MI 48060
810-982-7777 (P)
810-982-7794 (F)

BP060713 01